

AIA-360 Assay Specifications

ST CORT Test Code 067

No.	Item	Data
1	Calib. Req	Show
2	Cal 1	0.0 µg/dL
3	Cal 2	0.6 µg/dL (example)
4	Cal 3	2.0 µg/dL (example)
5	Cal 4	6.0 µg/dL (example)
6	Cal 5	20.0 µg/dL (example)
7	Cal 6	60.0 µg/dL (example)
8	Cal Lot L	
9	Cal Lot R	
10	Unit	µg/dL
11	Smpl. Vol	10
12	Dil. Vol	140
13	Assay L	0.2
14	Assay H	60
15	Ref. L	
16	Ref. H	
17	Decimal	1

Visible only in Test Mode

18	Test Name	#CORT
19	Calib. No	6
20	Calib. Mul	3
21	Calib. Equ	6
22	Calib. CV	90
23	Assay Prtl	1
24	Factor1 A	1.000000
25	Factor1 B	0.000000
26	Factor2 A	1.000000
27	Factor2 B	0.000000
28	V. Conc	0.000000
29	G Origin	0.000000

AIA-900 Assay Specifications

ST CORT Test Code 067

No.	Item	Data
1	Code	67
2	ACT	1
3	Analyte	#CORT
4	Lot	***Enter current cal lot no.
5	CAL 1	0.0 µg/dL
6	CAL 2	0.6 µg/dL (example)
7	CAL 3	2.0 µg/dL (example)
8	CAL 4	6.0 µg/dL (example)
9	CAL 5	20.0 µg/dL (example)
10	CAL 6	60.0 µg/dL (example)
11	Cal lot L	
12	Cal lot R	
13	Unit	ug/dL
14	Decimal	2
15	Assay Low	0.2
16	Assay High	60
17	Reference Low	
18	Reference High	
19	Reschedule Low	0.2
20	Reschedule High	60
21	Factor A	1
22	Factor B	0
23	Sample Volume	10
24	Diluent Volume	140
25	2 Step reagent dispensing volume	0
26	Calibration code	23
27	CAL. No.	6
28	CAL. MUL.	3
29	CAL. EQU.	6
30	CAL. CV	90
31	DIL. SP1	1
32	DIL. SP2	1
33	DIL. CAL. (Calculation of dil ratio of conc.)	1
34	DIL. CNTL.	1
35	DIL. DO	1
36	DIL. AH.	5
37	DIL. CALC.	1
38	DIL. CODE	23
39	DIL. NAME	CORT
40	DIL. PRTY	3
41	PRE. SPVOL (Vol. of pretreated sample)	0
42	PRE. 1VOL (Vol. of pretreatment sol-1)	0
43	PRE. 2 VOL (Vol. of pretreatment sol-2)	0
44	PRE. CODE (pretreated sol. code)	0
45	PRE. NAME (pretreated sol. name)	0
46	Protocol	1
47	SYS. F_A	1
48	SYS. F_B	0
49	V. CONC.	0
50	G. ORIGIN	0

AIA-2000 Assay Specifications

ST CORT Test Code 067

No.	Item	Data
1	Unit	µg/dL
2	Decimal places	1
3	Reference low	
4	Reference high	
5	Reschedule low	0.2
6	Reschedule high	60
7	Assay range low	0.2
8	Assay range high	60
9	Specimen diluent code	23
10	Specimen diluent name	CORT
11	Dilution factor for Sp. 1	1
12	Dilution factor for Sp. 2	1
13	Dilution factor for Control	1
14	Default multiplier for DO	3
15	Default multiplier for >H	3
16	Factor A	1.00000e+000
17	Factor B	0.00000e+000
18	Calibration code	6
19	Calibrator replicates	3
20	Calibrator lot	***Enter current cal. lot no.
21	Calibrator Concentration	
22	Cal - 01	0.0 µg/dL
23	Cal - 02	0.6 µg/dL (example)
24	Cal - 03	2.0 µg/dL (example)
25	Cal - 04	6.0 µg/dL (example)
26	Cal - 05	20.0 µg/dL (example)
27	Cal - 06	60.0 µg/dL (example)
28		
29		
30		
31		
32		
33		
34	Dilution factor for calibrator	1
35	Calibration period	90
36	Incubation time (10/40)	10
37	Assay protocol	1
38	Specimen volume	10
39	Diluent volume	140
40	Conjugate volume	0
41	Diluted conjugate volume	0
42	Pretreatment	0
43	Pretreatment specimen volume	0
44	Pretreatment Reagent code	0
45	Pretreatment Reagent name	0
46	Pretreatment Reagent 1 volume	0
47	Pretreatment Reagent 2 volume	0
48	System Factor	1
49	Calibration Code Check	023
50	Virtual Concentration	0.0
51	Graph Origin	0.0
52	Calculation with Dilution Factor	Yes

Cortisol

ST AIA-PACK CORT

Name and Intended Use

ST AIA-PACK CORT is designed for IN VITRO DIAGNOSTIC USE ONLY for the quantitative measurement of Cortisol in human serum or heparinized plasma on specific Tosoh AIA System analyzers.

Summary and Explanation of Test

Cortisol, the major glucocorticoid produced by the adrenal cortex regulates glucose, lipid and protein metabolism and has anti-inflammatory and immunosuppressive effects.^{1,2} The cortisol level in circulation is regulated by adrenocorticotrophic hormone (ACTH).³ ACTH is produced by the pituitary gland and is regulated by the hypothalamus with corticotropin releasing factor(s) (CRF).¹ The cortisol itself also negatively regulates ACTH production. Therefore, measurement of serum cortisol and monitoring a circadian pattern^{4,5} can provide information to evaluate hypothalamic-pituitary-adrenal function as well as clinical diagnosis.^{6,7}

Principle of the Assay

The ST AIA-PACK CORT is a competitive enzyme immunoassay which is performed entirely within the AIA-PACK. Cortisol present in the test sample competes with enzyme-labeled cortisol for a limited number of binding sites on a cortisol-specific antibody immobilized on magnetic beads. The beads are washed to remove the unbound enzyme labeled cortisol and are then incubated with a fluorogenic substrate, 4-methylumbelliferyl phosphate (4MUP). The amount of enzyme labeled cortisol that binds to the beads is inversely proportional to the cortisol concentration in the test sample. A standard curve using a range of known standard concentrations is constructed and unknown cortisol concentrations are calculated using this curve.

Material Provided (ST AIA-PACK CORT, Cat. No. 0025287)

5 trays x 20 test cups (ST AIA-PACK CORT Test Cup)

Plastic test cups containing lyophilized magnetic beads with anti-cortisol rabbit polyclonal antibody and cortisol conjugated to bovine alkaline phosphatase with 0.1% sodium azide as a preservative.

Materials Required But Not Provided

The following materials are not provided but are required to perform Cortisol analysis using the ST AIA-PACK CORT (Cat. No. 0025287) on specific Tosoh AIA Systems. They are available separately from Tosoh.

Materials	Cat. No.
AIA-SYSTEMS:	
AIA-360	0019945
AIA-900	0022930
AIA-900 9tray Sorter	0022931
AIA-900 19tray Sorter	0022932
AIA-2000 ST	0022100
AIA-2000 LA	0022101
AIA-PACK:	
AIA-PACK Substrate Set II	0020968
AIA-PACK Substrate/Reconstituent	
AIA-PACK CORT Calibrator Set	0020387
Calibrator #1 0.0 µg/dL	
Calibrator #2 0.6 µg/dL (approx.)	
Calibrator #3 2.0 µg/dL (approx.)	
Calibrator #4 6.0 µg/dL (approx.)	
Calibrator #5 20.0 µg/dL (approx.)	
Calibrator #6 60.0 µg/dL (approx.)	
AIA-PACK CORT Sample Diluting Solution	0020587
AIA-PACK Wash Concentrate Set	0020955
AIA-PACK Diluent Concentrate Set	0020956
AIA-PACK Detector Standardization Test Cups	0020970
AIA-PACK Sample Treatment Cups	0020971
Sample Cups	0018581
ADDITIONAL REQUIREMENTS: (Except AIA-360)	
Pipette Tips (1000/Pkg)	019215
Tip Rack (Empty)	019216
Preloaded Pipette Tips (96 Tips X 50 Racks)	996010
Preloaded Pipette Tips (96 Tips X 5 Racks)	996005

Only materials obtained from Tosoh should be used. Materials obtained elsewhere should not be substituted since assay performance is based strictly on Tosoh materials.

Warnings and Precautions

- The ST AIA-PACK CORT is intended for in vitro diagnostic use only.
- **Rx** only. US Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner.
- Inspect the packaging and the exterior of the aluminum pouch or the vials for any sign of damage before use. If any damages are visible, contact your local Tosoh sales representative.
- Test cups from different lots should not be mixed within a tray.
- The ST AIA-PACK CORT contains sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Human sera is not used in the preparation of this product, however, since human specimens will be used for samples and other quality control products in the lab may be derived from human serum, please use standard laboratory safety procedures in handling all specimens and controls.
- For safe waste disposal, it is recommended that each laboratory complies with established laboratory procedures and local, state, and federal regulations.
- Do not use beyond the expiration date.
- Serum, dust, metal, or microorganism contamination may cause degradation of reconstituted substrate solution. Store in a clean environment, away from direct sunlight and ultraviolet light.
- **Tosoh Automated Immunoassays utilizing alkaline phosphatase-based technologies should not be used with samples from patients under Asfotase Alfa treatment.**

Storage and Stability

All unopened materials are stable until the expiration date on the label when stored at the specified temperature.

Materials	Cat. No.
Refrigerator Temperature (2 - 8 °C):	
ST AIA-PACK CORT	0025287
AIA-PACK CORT Calibrator Set	0020387
AIA-PACK CORT Sample Diluting Solution	0020587
AIA-PACK Substrate Set II	0020968
AIA-PACK Wash Concentrate	0020955
AIA-PACK Diluent Concentrate	0020956
Room Temperature (18 - 25 °C):	
AIA-PACK Detector Standardization Test Cups	0020970
AIA-PACK Sample Treatment Cups	0020971

ST AIA-PACK CORT test cups may be stored for up to 24 hours at a room temperature of 18 - 25 °C. Calibrators must be kept tightly sealed and refrigerated at 2 - 8 °C. After opening, calibrators should be used within 7 days. After opening, Sample Diluting Solution is stable for up to 90 days refrigerated at 2 - 8 °C. Reconstituted substrate solution is stable for 3 days at 18 - 25 °C or 30 days at 2 - 8 °C. Working diluent and wash solutions are stable for 30 days at room temperature (18 - 25 °C). Reagents should not be used if they appear cloudy or discolored.

Specimen Collection and Handling

Serum or heparinized plasma is required for the assay. EDTA and citrated plasma **SHOULD NOT BE USED**.

Collection times should be standardized, usually 8 A.M. and 4 P.M. or 11 P.M. Timed, paired specimen collection is required for evaluation of baseline diurnal variation. Because of the normal diurnal variation and the circadian rhythm of cortisol secretion, the time of specimen collection should always be noted.⁴

No special patient preparation is necessary. When using serum, a venous blood sample is collected aseptically without additives. Store at 18 - 25 °C until a clot has formed (usually 15 - 45 minutes), then centrifuge to obtain the serum specimen for assay.

To use heparinized plasma, a venous blood sample is collected aseptically with the designated additive. Centrifuge and separate plasma from the packed cells as soon as possible.

Samples may be stored at 2 - 8 °C for up to 24 hours prior to analysis. If the analysis cannot be done within 24 hours, the sample should be stored frozen at -20 °C or below for up to 60 days.

Repeated freeze-thaw cycles should be avoided. Turbid serum samples or samples containing particulate matter should be centrifuged prior to testing. Prior to assay, slowly bring frozen samples to room temperature (18 - 25 °C) and mix gently.

The sample required for analysis is 10 µL.

Procedure

1) Reagent Preparation

1a) Substrate Solution

Bring all reagents to room temperature (18 - 25 °C) before preparing the working reagent. Add the entire contents of the Substrate Reconstituent (100 mL) to the lyophilized Substrate and mix thoroughly to dissolve the solid material.

1b) Wash Solution

Add the entire contents of the Wash Concentrate (100 mL) to approximately 2.0 L CAP Class I water or the clinical laboratory reagent water, mix well, and adjust the final volume to 2.5 L.

1c) Diluent

Add the entire contents of the Diluent Concentrate (100 mL) to approximately 4.0 L CAP Class I water or the clinical laboratory reagent water, mix well, and adjust the final volume to 5.0 L.

2) Calibration


2a) Calibration Curve

The calibrators for use with the ST AIA-PACK CORT are prepared gravimetrically and are compared to internal reference standards.

The calibration curve for the ST AIA-PACK CORT is stable for up to 90 days. Calibration stability is monitored by quality control performance and is dependent on proper reagent handling and AIA System maintenance according to the manufacturer's instructions.

Recalibration may be necessary more frequently if controls are out of the established range for this assay or if certain service procedures are performed (e.g. temperature adjustment, sampling mechanism changes, or detector lamp adjustment or change). For further information regarding instrument operation, consult the AIA System Operator's Manual.

The sample calibration curve below shows the algorithm used to calculate the results. This is an example only. Actual results will vary depending on the Instrument type and lot number used.



2b) Calibration Procedure

Refer to the appropriate AIA System Operator's Manual for procedural instructions.

- i) Verify that both the calibrator lot and concentration numbers have been correctly entered into the software.
- ii) Calibrators for ST AIA-PACK CORT are lyophilized and require reconstitution prior to use. Lyophilized calibrators should be reconstituted with 1.0 mL of CAP Class I water or the clinical laboratory reagent water. Tosoh recommends that all calibrators be run in triplicate.

2c) Calibration Acceptability criteria

- i) Since there is an inverse relationship between concentration and rate, the rates should decrease as the concentration increases.
- ii) The replicate values should be within a 10% range.

2d) Calibration Review and Acceptance

- i) Review the calibration curve carefully, using the criteria listed above.
- ii) Edit the calibration if necessary, then accept the calibration.
- iii) For further information regarding calibration, consult the specific AIA System Operator's Manual.

3) Quality Control

3a) Commercially Available Controls

Commercially available controls should be run at least once per day. It is recommended that at least two (2) levels of controls, normal and abnormal, be used. Laboratory policy for this particular assay designates the following:

Control Material: _____
Frequency: _____

Lot number of control material, acceptable limits, and corrective action to be taken if controls do not meet laboratory criteria will be found in a separate quality control document maintained by the laboratory.

3b) Quality Control Procedure

- i) Assay quality control specimens as instructed in the specific Operator's Manual for your analyzer. In addition, refer to the AIA System Operator's Manual for detailed instructions on defining and editing the files.
- ii) Quality control material to be run with this assay is defined by individual laboratory policy.

4) Specimen Processing

4a) Preparation

Following specific instructions in the Operator's Manual for the analyzer, place samples on the instrument appropriately.

4b) Assay Procedure

- i) Ensure a sufficient quantity of ST AIA-PACK CORT test cups for the number of samples to be run.
- ii) Load patient samples as instructed in the Operator's Manual and proceed with analysis. Note: The AIA-900 and AIA-2000 will require AIA-PACK Sample Treatment Cups if onboard dilutions are utilized.

Procedural Notes

- 1) Lyophilized Substrate must be completely dissolved.
- 2) Ligand assays performed by the Tosoh AIA System Analyzer require that the laboratory use water designated by the CAP Class I or by the clinical laboratory reagent water. Water should be tested at least once per month and should be free of particulate matter including bacteria. The pH of the water should also be routinely tested. For further information, consult the CLSI document GP40-A4-AMD, Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline - Fourth Edition.
- 3) If a specimen cortisol concentration is found to be greater than the linearity limit of the assay, 50 µg/dL; the specimen should be diluted with the AIA-PACK CORT Sample Diluting Solution and re-assayed according to the Assay Procedure. The recommended dilution for samples containing greater than the concentration of the highest calibrator is 1:3. It is desirable to dilute the sample so that the diluted sample reads between 10 and 50 µg/dL. The dilution factor should be entered into the software. For further information on the dilution of specimens, refer to the AIA System Operator's Manual.
- 4) The AIA systems can store two different calibration curves for each analyte at one time. Therefore, up to two different lots of ST AIA-PACK CORT Test cups can be used during the same run.
- 5) If the assay specifications for this test are not already in the system software, the specifications must be entered under test code 067.

Calculation of Results

The AIA Systems perform all sample and reagent handling operations automatically. The AIA Systems read the rate of fluorescence produced by the reaction and automatically convert the rate to cortisol concentration in µg/dL.

For samples requiring dilution, the AIA 900 and AIA-2000 will automatically perform dilutions and calculate results if the dilution factors are entered into the software. For detailed information regarding programming dilutions, consult the appropriate Operator's Manual.

Evaluation of Results

Quality Control

In order to monitor and evaluate the precision of the analytical performance, it is recommended that commercially available control samples be assayed daily.

The minimum recommendations for the frequency of running internal control material are:

- After calibration, two levels of controls are run in order to accept the calibration curve.
- The two levels of controls are also repeated after calibration when certain service procedures are performed (e.g. temperature adjustment, sampling mechanism changes, maintenance of the wash probe or detector lamp adjustment or change).
- After daily maintenance, at least two levels of the control should be run in order to verify the overall performance of the Tosoh AIA System Analyzers.

If one or more control sample value(s) is out of the acceptable range, it will be necessary to investigate the validity of the calibration curve before reporting patient results.

Standard laboratory procedures should be followed in accordance with the regulatory agency under which the laboratory operates.

Limitations of the Procedure

For diagnostic purposes, the results obtained from this assay should be used in conjunction with other data (e.g. symptoms, results of other tests, clinical impressions, therapy, etc.).

Using ST AIA-PACK CORT, the highest concentration of cortisol measurable without dilution is approximately 60 µg/dL, and the lowest measurable concentration is 0.2 µg/dL (assay sensitivity).

The exact linearity of the ST AIA-PACK CORT depends on the particular lot of calibrator in use. Although the approximate value of the highest calibrator is 60 µg/dL, the exact concentration may be slightly different. The assay specification, Assay Range High, should be defined as the upper limit of the assay range, 60 µg/dL.

Although hemolysis has an insignificant effect on the assay, hemolyzed samples may indicate mistreatment of a specimen prior to assay and results should be interpreted with caution.

Lipemia has an insignificant effect on the assay except in the case of gross lipemia where spatial interference may occur.

Certain medications may interfere with assay performance. Specimens from patients taking medicines and/or medical treatment may show erroneous results. All results should be interpreted with respect to the clinical picture of the patient.

For a more complete understanding of the limitations of this procedure, please refer to the Specimen Collection and Handling, Warnings and Precautions, Storage and Stability, and Procedural Notes sections in this insert sheet.

Expected Values

Each laboratory should determine a reference interval which corresponds to the characteristics of the population being tested. As with all diagnostic procedures, clinical results must be interpreted with regard to concomitant medications administered to the patient.⁸

Reference Ranges

Because of the diurnal variation and circadian rhythm of cortisol secretion, specimens drawn at 8 A.M. should represent peak levels in normal individuals. Specimens drawn at 11 P.M. will typically be 25% of the peak level.⁶

The interval given here was determined in serum samples which were collected at 8~9 a.m. from 266 apparently healthy individuals.

$$\text{Reference Interval} = 10.4 - 26.4 \mu\text{g/dL} \quad (287 - 729 \text{ nmol/L})$$

Conversion Factors

Cortisol concentrations in this application are in units of $\mu\text{g/dL}$. Conversion to SI units of nmol/L may be made using the following equations:

$$\text{nmol cortisol/L} = \mu\text{g cortisol/dL} \times 27.6$$

Performance Characteristics

The following performance characteristics were determined using the Tosoh AIA NexIA.

1) Accuracy

1a) Recovery:

Three serum pools were spiked with three different levels of cortisol and assayed before and after spiking.

Sample	Initial Value ($\mu\text{g/dL}$)	CORT Added ($\mu\text{g/dL}$)	Expected Value ($\mu\text{g/dL}$)	Measured Value ($\mu\text{g/dL}$)	Percent Recovery (%)
Serum A1	8.37	7.85	16.22	15.53	95.7
	8.37	15.69	24.06	24.19	100.5
	8.37	31.38	39.75	38.94	98.0
Serum B1	15.30	7.85	23.15	23.05	99.6
	15.30	15.69	30.99	31.43	101.4
	15.30	31.38	46.68	49.33	105.7
Serum C1	10.94	7.85	18.79	18.67	99.4
	10.94	15.69	26.63	26.52	99.6
	10.94	31.38	42.32	42.94	101.5

1b) Dilution:

Three serum samples containing high concentrations of cortisol were serially diluted with AIA-PACK CORT Sample Diluting Solution and assayed.

Sample	Dilution Factor	Expected Value (µg/dL)	Measured Value (µg/dL)	Percent Recovery (%)
Serum A2	none		47.25	
	7.5/10	35.44	36.89	104.1
	5.0/10	23.63	25.25	106.9
	2.5/10	11.81	12.25	103.7
	1.0/10	4.73	4.85	102.5
Serum B2	none		50.32	
	7.5/10	37.74	38.08	100.9
	5.0/10	25.16	25.48	101.3
	2.5/10	12.58	12.36	98.3
	1.0/10	5.03	4.83	96.0
Serum C2	none		52.18	
	7.5/10	39.13	41.49	106.0
	5.0/10	26.09	27.71	106.2
	2.5/10	13.04	13.28	101.8
	1.0/10	5.22	5.21	99.8

2) Precision

2a) Intra-assay precision

The intra-assay (within run) precision was determined using three controls in a total of 20 runs. Within each run, one set of duplicates per control was assayed. The mean of each duplicate was used to obtain the pooled standard deviation (SD), which was then used to calculate the coefficient of variation.

Sample	Mean (µg/dL)	Pooled SD (µg/dL)	Coefficient of Variation (%)
Serum A3	5.32	0.17	3.1
Serum B3	21.69	0.56	2.6
Serum C3	41.16	1.04	2.5

2b) Inter-assay precision

The inter-assay (between run) precision coefficient of variation was evaluated at three different concentrations by analyzing samples in 20 separate runs over 20 days.

Sample	Number of Replicates	Mean (µg/dL)	Standard Deviation (µg/dL)	Coefficient of Variation (%)
Serum A	20	5.30	0.25	4.7
Serum B	20	21.80	1.10	5.0
Serum C	20	41.04	2.11	5.1

2c)Total precision

Total precision was determined by the duplicate assay of three controls in 20 separate runs. The means of each run were used to calculate the standard deviation (SD) and coefficient of variation (CV).

Sample	Mean (µg/dL)	Standard Deviation (µg/dL)	Coefficient of Variation (%)
Serum A3	5.32	0.21	3.9
Serum B3	21.69	0.93	4.3
Serum C3	41.16	1.90	4.6

Specificity

The following substances were tested for cross-reactivity. The cross-reactivity (%) is the percent of the compound which will be identified as cortisol. If these compounds are present in the specimen at the same concentration as cortisol, the final result will be increased by these percentages.

Compound	Cross-reactivity (%)
Prednisolone	47.5
Cortisone	34.3
Prednisone	2.7
Corticosterone	2.6
17 α -Hydroxyprogesterone	0.28
Testosterone	<0.001
Progesterone	<0.001
17 β -estradiol	<0.001

Sensitivity

The minimal detectable concentration (MDC) of cortisol is estimated to be 0.2 µg/dL. The MDC is defined as that concentration of cortisol which corresponds to the rate of fluorescence that is two standard deviations from the mean rate of fluorescence of 20 replicate determinations of a zero calibrator.

Interference

Interference is defined, for purposes of this study, to be recovery outside of 10% of the known specimen mean concentration.

- Added hemoglobin (up to 430 mg/dL), free bilirubin (up to 17 mg/dL) and conjugated bilirubin (up to 19 mg/dL) do not interfere with the assay.
- Lipemia, as indicated by triglyceride concentrations (up to 1,660 mg/dL), does not interfere with the assay.
- Protein, as indicated by added albumin (up to 5 g/dL), does not interfere with the assay.
- Ascorbic acid (up to 20.0 mg/dL) does not interfere with the assay.
- Tosoh Automated Immunoassays utilizing alkaline phosphatase-based technologies should not be used with samples from patients under Asfotase Alfa treatment.

References

- 1) Baxter, J.D., Rousseau, G.G. 1979. Glucocorticoid Hormone Action: An Overview. In: Baxter, J.D. and Rousseau, G.G. (eds.) Glucocorticoid Hormone Action. Springer-Verlag, Berlin and New York.
- 2) Dallman, M.F., Yates, F.E. 1969. Dynamic Asymmetries in The Corticosteroid Feedback Path and Distribution-Metabolism-Binding Elements of The Adrenocortical System. Annals of New York Academy of Sciences. 156:696-721.
- 3) Keller-Wood, M.E. and Dallman, M.F., 1984, Corticosteroid Inhibition of ACTH Secretion. Endocrine Reviews, 5:1-24.
- 4) Weitzman, E.D., Schaumburg, H. and Fishbein, W., 1966, Plasma 17-Hydroxycorticosteroid Levels During Sleep in Man. J. Clin. Endocr., 26:121-127.
- 5) Orth, D.N. and Island, D.P., 1969, Light Synchronization of Circadian Rhythm in Plasma Cortisol (17-OHCS) Concentration in Man. J. Clin. Endocr., 29:479-486.
- 6) Vermeulen, A. and Van der Straeten, M., 1964, Determination of Plasma Cortisol by a Fluorometric Method. J. Clin. Endocr., 24:1188-1194.
- 7) Hubl, W., et al, 1984, A Sensitive Direct Enzyme Immunoassay for Cortisol in Plasma and Saliva. Exp. Clin. Endocrinol., 84:63-70.
- 8) Young, D., 1990, Effects of Drugs on Clinical Laboratory Tests. 3rd Edition, Washington, DC, American Association for Clinical Chemistry Press.

**TOSOH CORPORATION****Bioscience Division**

2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan

Phone: +81 3 6636 3734

Fax: +81 3 6636 3627

Website: www.tosoh.com

Supplied by**TOSOH BIOSCIENCE, INC.**

3600 Gantz Road,

Grove City, OH 43123 USA

Phone: +1 650 615 4970

Fax: +1 650 615 0415

Phone: (800) 248 6764

Fax: (800) 685 7595

Website: www.tosohbioscience.us



European Conformity



In vitro diagnostic medical device



Consult instructions for use



Temperature limitation



Batch code / Lot number



Manufacturer

Authorized representative
in the European Community

Use by date

Catalogue number
/ Part number

Supplied by



Sufficient for

US: For prescription use only

AIA-PACK CORT Calibrator Set

Intended Use

The AIA-PACK CORT Calibrator Set is intended for IN VITRO DIAGNOSTIC USE ONLY for the calibration of the ST AIA-PACK CORT Assays.

Summary and Explanation

The AIA-PACK CORT Calibrator Set contains human serum with assigned levels of cortisol (CORT). Calibration should be performed according to the schedule indicated in the AIA System Operator's Manual. The calibrators in this set are prepared gravimetrically and are compared to internal reference standards.

Material Provided (Cat. No. 0020387)

2 x 1 mL	Calibrator No. 1	0.0	µg/dL
	Human serum containing no detectable concentration of cortisol (0 µg cortisol/dL), and 0.1% sodium azide as a preservative. (Lyophilized)		

2 x 1 mL	Calibrator No. 2	0.6	µg/dL (approx.)
	No. 3	2.0	µg/dL (approx.)
	No. 4	6.0	µg/dL (approx.)
	No. 5	20.0	µg/dL (approx.)
	No. 6	60.0	µg/dL (approx.)

Human serum containing the assigned concentration of cortisol (described on each vial), and 0.1% sodium azide as a preservative. (Lyophilized)

Warnings and Precautions

- The AIA-PACK CORT Calibrator Set is for in vitro diagnostic use.
- These materials contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Human sera used in the preparation of this product has been tested by FDA cleared methods and found negative for the presence of HBsAg and antibody to HIV-1 and HCV. Because no test method can offer complete assurance that products derived from human blood will not transmit infectious agents, it is recommended that this product be handled with the same precautions as used for patient samples.
- Do not use beyond the expiration date.

Preparation and Storage

- Using a volumetric pipette or a calibrated adjustable pipette, reconstitute the lyophilized calibrators accurately with 1.0 mL of CAP Class I water or the clinical laboratory reagent water. Allow the lyophilized material to fully dissolve, then mix the calibrators gently but thoroughly prior to performing the calibration.
- Bring calibrator to room temperature (18 - 25 °C) for use.
- Always store the Calibrator Set in an upright position at 2 - 8 °C when not in use.

Stability

When stored unopened and refrigerated at 2 - 8 °C, the AIA-PACK CORT Calibrator Set is stable until the expiration date on the label. After opening, the calibrators should be used within 7 days.

Procedure

Refer to the CALIBRATION PROCEDURE in the AIA-PACK section of this analyte instructions. For additional procedural instructions regarding calibration, refer to the AIA System Operator's Manual.

Calibration

1. Verify that both the calibrator lot and concentrations have been correctly entered into the analyzer software.
2. Add the appropriate amounts of each calibrator to sample cups (refer to the instrument worklist for the amount required in each sample cup).
3. Place the sample cups and the test cups on the analyzer as indicated on the worklist.
4. Tosoh recommends that all calibrators be run in triplicate.

Assignment of Values

The AIA-PACK CORT Calibrator Set contains assigned concentrations of cortisol. The assigned value is determined on a lot-to-lot basis and is designed to provide an assay calibration range of approximately 0 to 60 µg/dL of cortisol. The Calibrator Set has been prepared gravimetrically and is compared to internal reference standards.

Results

Since there is an inverse relationship between concentration and rate, the rates should decrease as the concentration increases.

The replicate values should be within a 10% range.

Limitations

The AIA-PACK CORT Calibrator Set is designed solely for use with AIA-PACK assay procedures.

References

1. AIA Analyte Application Manual (AAM). Tosoh Bioscience, Inc., Grove City, OH.
2. AIA-System Operator's Manual. Tosoh Bioscience, Inc., Grove City, OH.



TOSOH CORPORATION
Bioscience Division
2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan
Phone: +81 3 6636 3734
Fax: +81 3 6636 3627
Website: www.tosoh.com

Supplied by

TOSOH BIOSCIENCE, INC.
3600 Gantz Road,
Grove City, OH 43123 USA
Phone: +1 650 615 4970
Fax: +1 650 615 0415
Phone: (800) 248 6764
Fax: (800) 685 7595
Website: www.tosohbioscience.us

 European Conformity	 In vitro diagnostic medical device	 Consult instructions for use	 Temperature limitation
 Batch code / Lot number	 Manufacturer	 Authorized representative in the European Community	 Use by date
 Catalogue number / Part number	 Supplied by	 Sufficient for	 US: For prescription use only

AIA-PACK

CORT Sample Diluting Solution

Intended Use

The AIA-PACK CORT Sample Diluting Solution is intended for IN VITRO DIAGNOSTIC USE ONLY to dilute patient samples that have concentrations of cortisol above the linear range of the assay.

Summary and Explanation

The AIA-PACK CORT Sample Diluting Solution contains human serum with no detectable concentration of Cortisol. This Sample Diluting Solution is to be used only with samples that are being tested for Cortisol concentrations using the ST AIA-PACK CORT Assays.

Material Provided (Cat. No. 0020587)

4 x 4 mL Sample Diluting Solution

Human serum containing no detectable concentration of cortisol (0 µg cortisol/dL), and 0.1% sodium azide as a preservative.

Warnings and Precautions

The AIA-PACK CORT Sample Diluting Solution is for in vitro diagnostic use.

These materials contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.

Human sera used in the preparation of this product has been tested by FDA cleared methods and found negative for the presence of HBsAg and antibody to HIV-1 and HCV. Because no test method can offer complete assurance that products derived from human blood will not transmit infectious agents, it is recommended that this product be handled with the same precautions as used for patient samples.

Do not use beyond the expiration date.

Preparation and Storage

The AIA-PACK CORT Sample Diluting Solution is provided ready to use.

Always store the Sample Diluting Solution in an upright position at 2 - 8 °C when not in use.

Stability

When stored unopened and refrigerated at 2 - 8 °C, the AIA-PACK CORT Sample Diluting Solution is stable until the expiration date on the label. After opening, the Sample Diluting Solution is stable for up to 90 days when refrigerated at 2 - 8 °C.

Procedure

Refer to the AIA System Operator's Manual for additional procedural instructions regarding sample dilution.

1. If a specimen is found to contain greater than the linearity limit of approximately 60.0 µg/dL, the specimen should be diluted with the Sample Diluting Solution and assayed according to the procedure in the AIA-PACK section of the analyte application.
2. The AIA-900 and AIA-2000 will perform dilutions automatically if the dilution factors are entered into the software prior to assaying the diluted sample.
3. The recommended dilution for serum containing greater than the concentration of the highest calibrator is 1:3. However, it is desirable to dilute the samples so that the diluted sample reads between 10 and 50 µg cortisol/dL.

Results

When an auto-dilution is performed, the AIA instrument will calculate the final result.

Limitations

The AIA-PACK CORT Sample Diluting Solution is designed solely for use with AIA-PACK assay procedures.

References

1. AIA Analyte Application Manual (AAM). Tosoh Bioscience, Inc., Grove City, OH.
2. AIA-System Operator's Manual Tosoh Bioscience, Inc., Grove City, OH.

**TOSOH CORPORATION****Bioscience Division**

2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan

Phone: +81 3 6636 3734

Fax: +81 3 6636 3627

Website: www.tosoh.com**Supplied by****TOSOH BIOSCIENCE, INC.**

3600 Gantz Road,

Grove City, OH 43123 USA

Phone: +1 650 615 4970

Fax: +1 650 615 0415

Phone: (800) 248 6764

Fax: (800) 685 7595

Website: www.tosohbioscience.us

European Conformity



In vitro diagnostic medical device



Consult instructions for use



Temperature limitation



Batch code / Lot number



Manufacturer

Authorized representative
in the European Community

Use by date

Catalogue number
/ Part number

Supplied by



Sufficient for

US: For prescription use only