

AIA-360 Assay Specifications

ST FER Test Code 064

No.	Item	Data
1	Calib. Req	Show
2	Cal 1	0 ng/mL
3	Cal 2	500 ng/mL (example)
4		
5		
6		
7		
8	Cal Lot L	
9	Cal Lot R	
10	Unit	ng/mL
11	Smpl. Vol	20
12	Dil. Vol	100
13	Assay L	3.0
14	Assay H	1000.0
15	Ref. L	
16	Ref. H	
17	Decimal	1

Visible only in Test Mode

18	Test Name	#FER
19	Calib. No	2
20	Calib. Mul	3
21	Calib. Equ	1
22	Calib. CV	90
23	Assay Prtl	1
24	Factor1 A	1.000000
25	Factor1 B	0.000000
26	Factor2 A	1.000000
27	Factor2 B	0.000000
28	V. Conc	0.000000
29	G Origin	0.000000

AIA-900 Assay Specifications

ST FER Test Code 064

No.	Item	Data
1	Code	64
2	ACT	1
3	Analyte	#FER
4	Lot	***Enter current cal lot no.
5	CAL 1	0 ng/mL
6	CAL 2	500 ng/mL (example)
7		
8		
9		
10		
11	Cal lot L	
12	Cal lot R	
13	Unit	ng/mL
14	Decimal	1
15	Assay Low	3.0
16	Assay High	1000
17	Reference Low	
18	Reference High	
19	Reschedule Low	3.0
20	Reschedule High	1000
21	Factor A	1
22	Factor B	0
23	Sample Volume	20
24	Diluent Volume	100
25	2 Step reagent dispensing volume	0
26	Calibration code	3
27	CAL. No.	2
28	CAL. MUL.	3
29	CAL. EQU.	1
30	CAL. CV	90
31	DIL. SP1	1
32	DIL. SP2	1
33	DIL. CAL. (Calculation of dil ratio of conc.)	1
34	DIL. CNTL.	1
35	DIL. DO	10
36	DIL. AH.	5
37	DIL. CALC.	1
38	DIL. CODE	3
39	DIL. NAME	FER
40	DIL. PRTY	3
41	PRE. SPVOL (Vol. of pretreated sample)	0
42	PRE. 1VOL (Vol. of pretreatment sol-1)	0
43	PRE. 2 VOL (Vol. of pretreatment sol-2)	0
44	PRE. CODE (pretreated sol. code)	0
45	PRE. NAME (pretreated sol. name)	0
46	Protocol	1
47	SYS. F_A	1
48	SYS. F_B	0
49	V. CONC.	0
50	G. ORIGIN	0

AIA-2000 Assay Specifications

ST FER Test Code 064

No.	Item	Data
1	Unit	ng/mL
2	Decimal places	1
3	Reference low	
4	Reference high	
5	Reschedule low	3.0
6	Reschedule high	1000.0
7	Assay range low	3.0
8	Assay range high	1000.0
9	Specimen diluent code	3
10	Specimen diluent name	FER
11	Dilution factor for Sp. 1	1
12	Dilution factor for Sp. 2	1
13	Dilution factor for Control	1
14	Default multiplier for DO	100
15	Default multiplier for >H	10
16	Factor A	1.00000e+000
17	Factor B	0.00000e+000
18	Calibration code	1
19	Calibrator replicates	3
20	Calibrator lot	***Enter current cal. lot no.
21	Calibrator Concentration	
22	Cal - 01	0 ng/mL
23	Cal - 02	500 ng/mL (example)
24		
25		
26		
27		
28		
29		
30		
31		
32		
33		
34	Dilution factor for calibrator	1
35	Calibration period	90
36	Incubation time (10/40)	10
37	Assay protocol	1
38	Specimen volume	20
39	Diluent volume	100
40	Conjugate volume	0
41	Diluted conjugate volume	0
42	Pretreatment	0
43	Pretreatment specimen volume	0
44	Pretreatment Reagent code	0
45	Pretreatment Reagent name	0
46	Pretreatment Reagent 1 volume	0
47	Pretreatment Reagent 2 volume	0
48	System Factor	1
49	Calibration Code Check	003
50	Virtual Concentration	0.0
51	Graph Origin	0.0
52	Calculation with Dilution Factor	Yes

Ferritin

ST AIA-PACK FER

Name and Intended Use

ST AIA-PACK FER is designed for IN VITRO DIAGNOSTIC USE ONLY for the quantitative measurement of Ferritin (FER) in human serum or heparinized plasma on Tosoh AIA System analyzers.

Summary and Explanation of Test

Ferritin is the major soluble iron storage protein from which iron may be mobilized for the synthesis of hemoglobin, myoglobin and other iron-containing proteins.¹⁻³ Ferritin is present in high concentrations in the cytoplasm of reticuloendothelial cells, liver cells, spleen cells and developing red cell precursors in bone marrow. Extracts from various tissues have different isoferritin distribution.³⁻⁶ The more acidic isoferritins are found primarily in heart, kidney and many tumor tissues. The liver, spleen and serum, on the other hand, contain the more basic isoferritins. Therefore, most immunoassays utilize basic isoferritins crystallized from human liver or spleen for the measurement of serum ferritin. Studies on patients with iron-deficiency and iron overload abnormalities confirmed that measurement of serum ferritin concentrations reflect the amount of storage iron in the body.⁴⁻¹²

Principle of the Assay

The ST AIA-PACK FER is a two-site immunoenzymometric assay which is performed entirely in the AIA-PACK. Ferritin present in the test sample is bound with monoclonal antibody immobilized on a magnetic solid phase and enzyme-labeled monoclonal antibody in the AIA-PACK. The magnetic beads are washed to remove unbound enzyme-labeled monoclonal antibody and are then incubated with a fluorogenic substrate, 4-methylumbelliferyl phosphate (4MUP). The amount of enzyme-labeled monoclonal antibody that binds to the beads is directly proportional to the ferritin concentration in the test sample. A standard curve is constructed, and unknown sample concentrations are calculated using this curve.

Material Provided (ST AIA-PACK FER, Cat. No. 0025253)

5 trays x 20 test cups (ST AIA-PACK FER Test Cup)

Plastic test cups containing lyophilized magnetic beads coated with anti-FER mouse monoclonal antibody and mouse monoclonal antibody (to human FER) conjugated to bovine alkaline phosphatase with 0.1% sodium azide as a preservative.

Materials Required But Not Provided

The following materials are not provided but are required to perform Ferritin analysis using the ST AIA-PACK FER (Cat. No. 0025253) on specific Tosoh AIA Systems. They are available separately from Tosoh.

Materials	Cat. No.
AIA-SYSTEMS:	
AIA-360	0019945
AIA-900	0022930
AIA-900 9tray Sorter	0022931
AIA-900 19tray Sorter	0022932
AIA-2000 ST	0022100
AIA-2000 LA	0022101
AIA-PACK:	
AIA-PACK Substrate Set II	0020968
AIA-PACK Substrate/Reconstituent	
AIA-PACK FER Calibrator Set	0020353
Calibrator Zero 0 ng/mL	
Positive 500 ng/mL (approx.)	
AIA-PACK FER Sample Diluting Solution	0020553
AIA-PACK Wash Concentrate Set	0020955
AIA-PACK Diluent Concentrate Set	0020956
AIA-PACK Detector Standardization Test Cups	0020970
AIA-PACK Sample Treatment Cups	0020971
Sample Cups	0018581
ADDITIONAL REQUIREMENTS: (Except AIA-360)	
Pipette Tips (1000/Pkg)	019215
Tip Rack (Empty)	019216
Preloaded Pipette Tips (96 Tips X 50 Racks)	996010
Preloaded Pipette Tips (96 Tips X 5 Racks)	996005

Only materials obtained from Tosoh should be used. Materials obtained elsewhere should not be substituted since assay performance is based strictly on Tosoh materials.

Warnings and Precautions

- The ST AIA-PACK FER is intended for in vitro diagnostic use only.
- **Rx** only. US Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner.
- Inspect the packaging and the exterior of the aluminum pouch or the vials for any sign of damage before use. If any damages are visible, contact your local Tosoh sales representative.
- Test cups from different lots should not be mixed within a tray.
- The ST AIA-PACK FER contains sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Human sera is not used in the preparation of this product, however, since human specimens will be used for samples and other quality control products in the lab may be derived from human serum, please use standard laboratory safety procedures in handling all specimens and controls.
- For safe waste disposal, it is recommended that each laboratory complies with established laboratory procedures and local, state, and federal regulations.
- Do not use beyond the expiration date.
- The ST AIA-PACK FER has been designed so that the high dose “hook effect” is not a problem for the vast majority of samples. Samples with ferritin concentrations between 1000 and 20,000 ng/mL will read > 1000 ng/mL. The “hook effect” phenomenon may occur only at ferritin concentrations > 10,000 ng/mL.
- Serum, dust, metal, or microorganism contamination may cause degradation of reconstituted substrate solution. Store in a clean environment, away from direct sunlight and ultraviolet light.
- Tosoh Automated Immunoassays utilizing alkaline phosphatase-based technologies should not be used with samples from patients under Asfotase Alfa treatment.

Storage and Stability

All unopened materials are stable until the expiration date on the label when stored at the specified temperature.

Materials	Cat. No.
Refrigerator Temperature (2 - 8 °C):	
ST AIA-PACK FER	0025253
AIA-PACK FER Calibrator Set	0020353
AIA-PACK FER Sample Diluting Solution	0020553
AIA-PACK Calibration Verification/Linearity Test Set	0020653
AIA-PACK Substrate Set II	0020968
AIA-PACK Wash Concentrate	0020955
AIA-PACK Diluent Concentrate	0020956

Room Temperature (18 - 25 °C):

AIA-PACK Detector Standardization Test Cups	0020970
AIA-PACK Sample Treatment Cups	0020971

ST AIA-PACK FER test cups may be stored for up to 24 hours at a room temperature of 18 - 25 °C. Calibrators must be kept tightly sealed and refrigerated at 2 - 8 °C. After opening, calibrators should be used within 24 hours. After opening, Sample Diluting Solution is stable for up to 90 days refrigerated at 2 - 8 °C. Reconstituted substrate solution is stable for 3 days at 18 - 25 °C or 30 days at 2 - 8 °C. Working diluent and wash solutions are stable for 30 days at room temperature (18 - 25 °C). Reagents should not be used if they appear cloudy or discolored.

Specimen Collection and Handling

Serum or heparinized plasma is required for the assay. EDTA and citrated plasma **SHOULD NOT BE USED**.

No special patient preparation is necessary. When using serum, a venous blood sample is collected aseptically without additives. Store at 18 - 25 °C until a clot has formed (usually 15 - 45 minutes), then centrifuge to obtain the serum specimen for assay. Capillary samples should not be used for ferritin determinations. Ferritin levels in capillary blood are significantly higher than ferritin levels in venous blood.¹³

To use heparinized plasma, a venous blood sample is collected aseptically with the designated additive. Centrifuge and separate plasma from the packed cells as soon as possible.

Samples may be stored at 2 - 8 °C for up to 24 hours prior to analysis. If the analysis cannot be done within 24 hours, the sample should be stored frozen at -20 °C or below for up to 60 days.

Repeated freeze-thaw cycles should be avoided. Turbid serum samples or samples containing particulate matter should be centrifuged prior to testing. Prior to assay, slowly bring frozen samples to room temperature (18 - 25 °C) and mix gently.

The sample required for analysis is 20 µL.

Procedure

1) Reagent Preparation

1a) Substrate Solution

Bring all reagents to room temperature (18 - 25 °C) before preparing the working reagent. Add the entire contents of the Substrate Reconstituent (100 mL) to the lyophilized Substrate and mix thoroughly to dissolve the solid material.

1b) Wash Solution

Add the entire contents of the Wash Concentrate (100 mL) to approximately 2.0 L of CAP Class I water or the clinical laboratory reagent water, mix well, and adjust the final volume to 2.5 L.

1c) Diluent

Add the entire contents of the Diluent Concentrate (100 mL) to approximately 4.0 L of CAP Class I water or the clinical laboratory reagent water, mix well, and adjust the final volume to 5.0 L.

2) Calibration


2a) Calibration Curve

The calibrators for use with the ST AIA-PACK FER have been standardized against WHO 1st IS 80/602 (1983).

The calibration curve for the ST AIA-PACK FER is stable for up to 90 days. Calibration stability is monitored by quality control performance and is dependent upon proper reagent handling and AIA System maintenance according to the manufacturer's instructions.

Recalibration may be necessary more frequently if controls are out of the established range for this assay or if certain service procedures are performed (e.g. temperature adjustment, sampling mechanism changes, or detector lamp adjustment or change). For further information regarding instrument operation, consult the AIA System Operator's Manual.

The sample calibration curve below shows the algorithm used to calculate the results. This is an example only. Actual results will vary depending on the Instrument type and lot number used.



2b) Calibration Procedure

Refer to the appropriate AIA System Operator's Manual for procedural instructions.

- Verify that both the calibrator lot and concentration numbers have been correctly entered into the software.
- Calibrators for ST AIA-PACK FER are provided ready to use. Tosoh recommends that all calibrators be run in triplicate.

2c) Calibration Acceptability criteria

- iii) The mean rate for the zero calibrator should be < 3.0 nM/sec.
- iv) Since there is a direct relationship between concentration and rate, the rates should increase as the concentration increases.
- v) The replicate values should be within a 10% range.

2d) Calibration Review and Acceptance

- i) Review the calibration curve carefully, using the criteria listed above.
- ii) Edit the calibration if necessary, then accept the calibration.
- iii) For further information regarding calibration, consult the specific AIA System Operator's Manual.

3) Quality Control

3a) Commercially Available Controls

Commercially available controls should be run at least once per day. It is recommended that at least two (2) levels of controls, normal and abnormal, be used. Laboratory policy for this particular assay designates the following:

Control Material: _____
Frequency: _____

Lot number of control material, acceptable limits, and corrective action to be taken if controls do not meet laboratory criteria will be found in a separate quality control document maintained by the laboratory.

3b) Quality Control Procedure

- i) Assay quality control specimens as instructed in the specific Operator's Manual for your analyzer. In addition, refer to the AIA System Operator's Manual for detailed instructions on defining and editing the files.
- ii) Quality control material to be run with this assay is defined by individual laboratory policy.

4) Specimen Processing

4a) Preparation

Following specific instructions in the Operator's Manual for the analyzer, place samples on the instrument appropriately.

4b) Assay Procedure

- i) Ensure a sufficient quantity of ST AIA-PACK FER test cups for the number of samples to be run.
- ii) Load patient samples as instructed in the Operator's Manual and proceed with analysis. Note: The AIA-900 and AIA-2000 will require AIA-PACK Sample Treatment Cups if onboard dilutions are utilized.

Procedural Notes

- 1) Lyophilized Substrate must be completely dissolved.
- 2) Ligand assays performed by the Tosoh AIA System Analyzer require that the laboratory use water designated by the CAP Class I or by the clinical laboratory reagent water. Water should be tested at least once per month and should be free of particulate matter including bacteria. The pH of the water should also be routinely tested. For further information, consult the CLSI document GP40-A4-AMD, Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline - Fourth Edition.
- 3) If a serum or plasma specimen Ferritin concentration is found to be greater than the linearity limit of the assay, 1000 ng/mL; the specimen should be diluted with the FER Sample Diluting Solution and re-assayed according to the Assay Procedure. The recommended dilution for samples containing greater than 1000 ng/mL is 1:10 or 1:100. It is desirable to dilute the sample so that the diluted sample reads between 50 and 500 ng/mL. The dilution factor should be entered into the software. For further information on the dilution of specimens, refer to the AIA System Operator's Manual.
- 4) The AIA systems can store two different calibration curves for each analyte at one time. Therefore, up to two different lots of ST AIA-PACK FER Test cups can be used during the same run.
- 5) If the assay specifications for this test are not already in the system software, the specifications must be entered under test code 064.

Calculation of Results

The AIA Systems perform all sample and reagent handling operations automatically. The AIA Systems read the rate of fluorescence produced by the reaction and automatically convert the rate to ferritin concentration in ng/mL.

For samples requiring dilution, the AIA-900 and AIA-2000 will automatically perform dilutions and calculate results if the dilution factors are entered into the software. For detailed information regarding programming dilutions, consult the appropriate Operator's Manual.

Evaluation of Results

Quality Control

In order to monitor and evaluate the precision of the analytical performance, it is recommended that commercially available control samples be assayed daily.

The minimum recommendations for the frequency of running internal control material are:

- After calibration, two levels of controls are run in order to accept the calibration curve.
- The two levels of controls are also repeated after calibration when certain service procedures are performed (e.g. temperature adjustment, sampling mechanism changes, maintenance of the wash probe or detector lamp adjustment or change).
- After daily maintenance, at least two levels of the control should be run in order to verify the overall performance of the Tosoh AIA System Analyzers.

If one or more control sample value(s) is out of the acceptable range, it will be necessary to investigate the validity of the calibration curve before reporting patient results.

Standard laboratory procedures should be followed in accordance with the regulatory agency under which the laboratory operates.

Limitations of the Procedure

For diagnostic purposes, the results obtained from this assay should be used in conjunction with other data (e.g., symptoms, results of other tests, clinical impressions, therapy, etc.).

Using ST AIA-PACK FER, the highest concentration of Ferritin measurable without dilution is 1000 ng/mL, and the lowest measurable concentration is 3.0 ng/mL (assay sensitivity).

Although the approximate value of the highest calibrator is 500 ng/mL, the exact concentration may be slightly different. The assay specification, Assay Range High, should be defined as the upper limit of the assay range, 1000 ng/mL.

Although hemolysis has an insignificant effect on the assay, hemolyzed samples may indicate mistreatment of a specimen prior to assay and results should be interpreted with caution.

Lipemia has an insignificant effect on the assay except in the case of gross lipemia where spatial interference may occur.

Specimens from patients who have received preparations of mouse monoclonal antibodies for diagnosis or therapy may contain human anti-mouse antibodies (HAMA). Such specimens may show falsely elevated values when tested for Ferritin.

Certain medications may interfere with assay performance. Specimens from patients taking medicines and/or medical treatment may show erroneous results. All results should be interpreted with respect to the clinical picture of the patient.

For a more complete understanding of the limitations of this procedure, please refer to the Specimen Collection and Handling, Warnings and Precautions, Storage and Stability, and Procedural Notes sections in this insert sheet.

Expected Values

Each laboratory should determine a reference interval which corresponds to the characteristics of the population being tested. As with all diagnostic procedures, clinical results must be interpreted with regard to concomitant medications administered to the patient.¹⁴

Reference Ranges

The intervals given here were determined in serum samples from apparently healthy individuals.

Age	Male	Female
18 to 45 yrs	22 - 340 ng/mL	6 - 115 ng/mL
> 45 yrs	22 - 415 ng/mL	15 - 200 ng/mL

Performance Characteristics

The following performance characteristics were determined using the Tosoh AIA NexIA. The AIA-600 II demonstrates equivalent performance.

1) Accuracy

1a) Recovery:

Three serum pools were spiked with three different levels of FER and assayed before and after spiking.

Sample	Initial Value (ng/mL)	FER Added (ng/mL)	Expected Value (ng/mL)	Measured Value (ng/mL)	Percent Recovery (%)
Serum A1	111.7	142.3	254.0	235.3	92.6
	111.7	284.6	396.3	368.0	92.9
	111.7	569.1	680.8	649.8	95.4
Serum B1	98.4	142.3	240.7	219.7	91.3
	98.4	284.6	383.0	347.9	90.8
	98.4	569.1	667.5	620.3	92.9
Serum C1	17.1	142.3	159.4	147.4	92.5
	17.1	284.6	301.7	280.3	92.8
	17.1	569.1	586.2	551.8	94.1

1b) Dilution:

Three serum samples containing high concentrations of ferritin were serially diluted with AIA-PACK FER Sample Diluting Solution and assayed.

Sample	Dilution Factor	Expected Value (ng/mL)	Measured Value (ng/mL)	Percent Recovery (%)
Serum A2	none		1083.8	
	7.5/10	812.9	779.8	95.9
	5.0/10	541.9	501.6	92.6
	2.5/10	271.0	235.0	86.7
	1.0/10	108.4	95.0	87.6
Serum B2	none		983.9	
	7.5/10	737.9	729.7	98.9
	5.0/10	491.9	474.9	96.5
	2.5/10	246.0	230.5	93.7
	1.0/10	98.4	87.3	88.7
Serum C2	none		933.8	
	7.5/10	700.4	713.4	101.9
	5.0/10	466.9	477.4	102.2
	2.5/10	233.5	238.2	102.0
	1.0/10	93.4	92.3	98.8

2) Precision**2a) Intra-assay precision**

The intra-assay (within run) precision was determined using three controls in a total of 20 runs. Within each run, one set of duplicates per control was assayed. The mean of each duplicate was used to obtain the pooled standard deviation (SD), which was then used to calculate the coefficient of variation.

Sample	Mean (ng/mL)	Pooled SD (ng/mL)	Coefficient of Variation (%)
Serum A3	20.5	0.48	2.3
Serum B3	246.6	5.57	2.3
Serum C3	673.4	15.39	2.3

2b) Inter-assay precision

The inter-assay (between run) precision coefficient of variation was evaluated at three different concentrations by analyzing samples in 20 separate runs over 20 days.

Sample	Number of Replicates	Mean (ng/mL)	Standard Deviation (ng/mL)	Coefficient of Variation (%)
Serum A3	20	20.0	0.74	3.7
Serum B3	20	240.0	7.61	3.2
Serum C3	20	657.6	32.36	4.9

2c) Total precision

Total precision was determined by the duplicate assay of three controls in 20 separate runs. The means of each run were used to calculate the standard deviation (SD) and coefficient of variation (CV).

Sample	Mean (ng/mL)	Standard Deviation (ng/mL)	Coefficient of Variation (%)
Serum A3	20.5	1.01	4.9
Serum B3	246.6	10.57	4.3
Serum C3	673.4	31.47	4.7

Specificity

The following substances were tested for cross-reactivity. The cross-reactivity (%) is the percent of the compound that will be identified as ferritin. If these compounds are present in the specimen at the same concentration as ferritin, the final result will be increased by these percentages.

Compound	Cross-reactivity (%)
Liver Ferritin	96
Spleen Ferritin	124
Heart Ferritin	6
Placental Ferritin	143

Sensitivity

The minimal detectable concentration (MDC) of Ferritin is estimated to be 3.0 ng/mL. The MDC is defined as that concentration of FER which corresponds to the rate of fluorescence that is two standard deviations from the mean rate of fluorescence of 20 replicate determinations of a zero calibrator.

Interference

Interference is defined, for purposes of this study, to be recovery outside of 10% of the known specimen mean concentration.

- Added hemoglobin (up to 430 mg/dL), free bilirubin (up to 17 mg/dL) and conjugated bilirubin (up to 19 mg/dL) do not interfere with the assay.
- Lipemia, as indicated by triglyceride concentrations (up to 1660 mg/dL), does not interfere with the assay.
- Ascorbic acid (up to 20 mg/dL) does not interfere with the assay.
- Protein, as indicated by human albumin concentrations (up to 5 g/dL), does not interfere with the assay.
- Tosoh Automated Immunoassays utilizing alkaline phosphatase-based technologies should not be used with samples from patients under Asfotase Alfa treatment.

References

- 1) Clichton, R.R., 1971, Ferritin: Structure, Synthesis and Function. *New Eng. J. Med.* 284:1413.
- 2) Clegg, G.A., Litton, J.E., Harrison, P.N., and Teffry, A., 1980, Ferritin: Molecular Structure and Iron Storage Mechanism. *Prog. Biophys. Molec. Biol.* 36:56.
- 3) Drysdale, J.W., Kohgo, Y., and Watanabe, N., 1980, Ferritin Phenotypes. In "Radioimmunoassay of Hormones, Proteins, and Enzymes," A. Albertini (ed.). *Excerpta Medica, Amsterdam*, p. 213.
- 4) Jacobs, A., Path, F.R.C., and Worwood, M., 1975, Ferritin in Serum: Clinical and Biochemical Implications. *New Eng. J. Med.*, 292:951.
- 5) Worwood, M., 1979, Serum Ferritin. In "CRC Critical Reviews in Clin. Lab. Sci." 171.
- 6) Halliday, J.W., Powell, L.W., 1979, Serum Ferritin and Isoferritins in Clinical Medicine. *Prog. Haematol.* 11:229.
- 7) Addison, G.M., Beamish, M.R., Hales, C.N., Hodgkins, M., Jacobs, A., and Liewellin, P., 1972, An immunoradiometric Assay for Ferritin in the Serum of Normal Subjects and Patients with Iron-Deficiency and Iron Overload. *J. Clin. Path.* 25:326.
- 8) Forman, D.T., and Parke, S.L., 1980, The Measurement and Interpretation of Serum Ferritin. *Ann. Clin. Lab. Sci.* 10:345.
- 9) Jacobs, A., Miller, F., Worwood, M., Beamish, M.R., and Wardrop, C.A., 1972, Ferritin in the Serum of Normal Subjects and Patients with Iron-Deficiency and Iron Overload. *Br. Med. J.* 4:206.
- 10) Walters, G.O., Miller, F.W., Worwood, M., 1973, Serum Ferritin Concentration and Iron Stores in Normal Subjects. *J. Clin. Path.* 26:7705.
- 11) Lipschitz, D.A., Cook, J.D., Finch, C.A., 1974, A Clinical Evaluation of Serum Ferritin as an Index of Iron Stores. *New Eng. J. Med.*, 290:1213.
- 12) Cook, J.D., Lipschitz, D.A., Miles, L.E.M., and Finch, C.A., 1974, Serum Ferritin as a Measure of Iron Stores in Normal Subjects. *Am. J. Clin. Nutr.* 27:681.

- 13) Mejia, L.A., and Viteri, F.E., 1983, Ferritin Concentrations in Plasma from Capillary (finger prick) Blood and Venous Blood Compared. Clin. Chem. 29: 871.
- 14) Young, D., 1990, Effects of Drugs on Clinical Laboratory Tests. 3rd Edition, Washington, DC, American Association for Clinical Chemistry Press.



TOSOH CORPORATION
Bioscience Division
 2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan
 Phone: +81 3 6636 3734
 Fax: +81 3 6636 3627
 Website: www.tosoh.com

Supplied by

TOSOH BIOSCIENCE, INC.
 3600 Gantz Road,
 Grove City, OH 43123 USA
 Phone: +1 650 615 4970
 Fax: +1 650 615 0415
 Phone: (800) 248 6764
 Fax: (800) 685 7595
 Website: www.tosohbioscience.us

 European Conformity	 <i>In vitro</i> diagnostic medical device	 Consult instructions for use	 Temperature limitation
 Batch code / Lot number	 Manufacturer	 Authorized representative in the European Community	 Use by date
 Catalogue number / Part number	 Supplied by	 Sufficient for	

AIA-PACK FER Calibrator Set

Intended Use

The AIA-PACK FER Calibrator Set is intended for IN VITRO diagnostic use only for the calibration of the ST AIA-PACK FER Assays.

Summary and Explanation

The AIA-PACK FER Calibrator Set contains human serum with assigned levels of Ferritin. Calibration should be performed according to the schedule indicated in the AIA System Operator's Manual. The calibrators in this set have been standardized against WHO 1st IS 80/602 (1983).

Material Provided (Cat. No. 0020353)

2 x 1 mL	Zero Calibrator
	Human serum containing no detectable concentration of ferritin (0 ng ferritin/mL), and 0.1% sodium azide as preservative. (ready to use)
2 x 1 mL	Positive Calibrator
	Human serum containing the assigned concentration of ferritin (approximately 500 ng ferritin/mL, described on each vial) and 0.1% sodium azide as preservative. (ready to use)

Warnings and Precautions

- The AIA-PACK FER Calibrator Set is for in vitro diagnostic use.
- These materials contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Human sera used in the preparation of this product has been tested by FDA cleared methods and found negative for the presence of HBsAg and antibody to HIV-1 and HCV. Because no test method can offer complete assurance that products derived from human blood will not transmit infectious agents, it is recommended that this product be handled with the same precautions as used for patient samples.
- Do not use beyond the expiration date.

Preparation and Storage

- The AIA-PACK FER calibrators are provided ready to use.
- Bring calibrator to room temperature (18 - 25 °C) prior to use.
- Always store the Calibrator Set in an upright position at 2 - 8 °C when not in use.

Stability

When stored unopened and refrigerated at 2 - 8 °C, the AIA-PACK FER Calibrator Set is stable until the expiration date on the label. After opening, the calibrators should be used within 24 hours.

Procedure

Refer to the CALIBRATION PROCEDURE in the AIA-PACK section of this analyte application. For additional procedural instructions regarding calibration, refer to the AIA System Operator's Manual.

- 1) Verify that both the calibrator lot and concentrations have been correctly entered into the analyzer software.
- 2) Add the appropriate amounts of each calibrator to sample cups (refer to the instrument worklist for the amount required in each sample cup).
- 3) Place the sample cups and the test cups on the analyzer as indicated on the worklist.
- 4) Tosoh recommends that all calibrators be run in triplicate.

Assignment of Values

The AIA-PACK FER Calibrator Set contains assigned concentrations of Ferritin. The assigned value is determined on a lot-to-lot basis and is designed to provide an assay calibration range of approximately 0.0 to 1000.0 ng/mL of Ferritin. The calibrators in this set have been standardized against WHO 1st IS 80/602 (1983).

Results

- The mean rate for the zero calibrator should be < 3.0 nM/sec.
- Since there is a direct relationship between concentration and rate, the rates should increase as the concentration increases.
- The replicate values should be within a 10% range.

Limitations

The AIA-PACK FER Calibrator Set is designed solely for use with AIA-PACK assay procedures.

References

- 1) AIA Analyte Application Manual (AAM). Tosoh Bioscience, Inc., Inc., Grove City, OH.
- 2) AIA-System Operator's Manual. Tosoh Bioscience, Inc., Grove City, OH.



TOSOH CORPORATION
Bioscience Division
2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan
Phone: +81 3 6636 3734
Fax: +81 3 6636 3627
Website: www.tosoh.com

Supplied by

TOSOH BIOSCIENCE, INC.
3600 Gantz Road,
Grove City, OH 43123 USA
Phone: +1 650 615 4970
Fax: +1 650 615 0415
Phone: (800) 248 6764
Fax: (800) 685 7595
Website: www.tosohbioscience.us



European Conformity



In vitro diagnostic medical device



Consult instructions for use



Temperature limitation



Batch code / Lot number



Manufacturer



Authorized representative
in the European Community



Use by date




Catalogue number
/ Part number



Supplied by



Sufficient for



US: For prescription use only

AIA-PACK

FER Sample Diluting Solution

Intended Use

The AIA-PACK FER Sample Diluting Solution is intended for IN VITRO DIAGNOSTIC USE ONLY to dilute patient samples that have concentrations of Ferritin above the linear range of the assay.

Summary and Explanation

The AIA-PACK FER Sample Diluting Solution contains human serum with no detectable concentration of Ferritin. This Sample Diluting Solution is to be used only with samples that are being tested for Ferritin concentrations using the ST AIA-PACK FER assays.

Materials Provided (Cat. No. 0020553)

4 x 4 mL Sample Diluting Solution
Human serum containing no detectable concentration of ferritin (0 ng ferritin/mL), and 0.1% sodium azide as a preservative.

Warnings and Precautions

- The AIA-PACK FER Sample Diluting Solution is for IN VITRO diagnostic use.
- These materials contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with large volumes of water to prevent azide build-up.
- Human sera used in the preparation of this product has been tested by FDA cleared methods and found negative for the presence of HBsAg and antibody to HIV-1 and HCV. Because no test method can offer complete assurance that products derived from human blood will not transmit infectious agents, it is recommended that this product be handled with the same precautions as used for patient samples.
- Do not use beyond the expiration date.

Preparation and Storage

- The AIA-PACK FER Sample Diluting Solution is provided ready to use.
- Always store the Sample Diluting Solution in an upright position at 2 - 8 °C when not in use.

Stability

When stored unopened and refrigerated at 2 - 8 °C, the AIA-PACK FER Sample Diluting Solution is stable until the expiration date on the label. After opening, the Sample Diluting Solution is stable for up to 90 days when refrigerated at 2 - 8 °C.

Procedure

Refer to the AIA System Operator's Manual for additional procedural instructions regarding sample dilution.

1. If a specimen is found to contain greater than the linearity limit of 1000 ng/mL, the specimen should be diluted with the Sample Diluting Solution and assayed according to the procedure in the AIA-PACK section of the analyte application.
2. The AIA-900 and AIA-2000 will perform dilutions automatically if the dilution factors are entered into the software prior to assaying the diluted sample.
3. The recommended dilution for specimen containing greater than 1000 ng/mL is 1:10 or 1:100. However, it is desirable to dilute the samples that contain more than 1000 ng ferritin/mL so that the diluted sample reads between 50 and 500 ng ferritin/mL.

Results

When an auto-dilution is performed, the AIA instrument will calculate the final result.

Limitations

The AIA-PACK FER Sample Diluting Solution is designed solely for use with AIA-PACK assay procedures.

References

- 1) AIA Analyte Application Manual (AAM). Tosoh Bioscience, Inc., Grove City, OH.
- 2) AIA-System Operator's Manual. Tosoh Bioscience, Inc., Grove City, OH.



TOSOH CORPORATION
Bioscience Division
2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan
Phone: +81 3 6636 3734
Fax: +81 3 6636 3627
Website: www.tosoh.com

Supplied by

TOSOH BIOSCIENCE, INC.
3600 Gantz Road,
Grove City, OH 43123 USA
Phone: +1 650 615 4970
Fax: +1 650 615 0415
Phone: (800) 248 6764
Fax: (800) 685 7595
Website: www.tosohbioscience.us



European Conformity



In vitro diagnostic medical device



Consult instructions for use



Temperature limitation



Batch code / Lot number



Manufacturer



Authorized representative
in the European Community



Use by date



Catalogue number
/ Part number



Supplied by



Sufficient for

US: For prescription use only

AIA-PACK

FER Calibration Verification/Linearity Test Set

Intended Use

The AIA-PACK FER Calibration Verification/Linearity Test Set is intended for IN VITRO diagnostic use only for the verification of the calibration and linearity of the ST AIA-PACK FER Assays.

Summary and Explanation

The AIA-PACK FER Calibration Verification/Linearity Test Set can be used as part of a quality assurance program to assist in complying with various regulatory requirements under which an institution may operate. Specific guidelines for the acceptability of data are determined by the institution in conformance with these requirements. Good laboratory practices stipulate using different materials for calibration and quality control purposes. Calibration Verification Materials should be analyzed as unknown test samples according to the directions stated in the AIA Systems Operator's Manual.

Materials Provided (Cat. No. 0020653)

2 x 4 mL	Sample Diluting Solution (SDS) Human serum containing no detectable concentration of ferritin (0 ng ferritin/mL), and 0.1% sodium azide as a preservative. (Ready to Use)
2 x 2 mL	Calibration Verification Material (CVM) Human serum containing the assigned concentration of ferritin (approximate upper linearity limit of 1000 ng ferritin/mL described on each vial) and 0.1% sodium azide as a preservative. (Ready to Use)

Warnings and Precautions

- The AIA-PACK FER Calibration Verification/Linearity Test Set is for in vitro diagnostic use.
- These materials contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such materials, always flush with large volumes of water to prevent azide build-up.
- Human sera used in the preparation of this product has been tested by FDA cleared methods and found negative for the presence of HBsAg and antibody to HIV-1 and HCV. Because no test method can offer complete assurance that products derived from human blood will not transmit infectious agents, it is recommended that this product be handled with the same precautions as used for patient samples.
- Do not use beyond the expiration date.

Preparation and Storage

- The Calibration Verification Material and Sample Diluting Solution are provided ready to use.
- The materials should be at room temperature (18 - 25 °C) prior to use.
- Always store the Calibration Verification/Linearity Test Set in an upright position at 2 - 8 °C when not in use.

Stability

When stored unopened and refrigerated at 2 - 8 °C, the AIA-PACK FER Calibration Verification/Linearity Test Set is stable until the expiration date on the label. After opening, the materials should be used within 24 hours.

Procedure

The frequency of performing calibration verification and linearity of the ST AIA-PACK FER assay is established by each institution in accordance with its quality assurance program and appropriate regulatory requirements. Refer to the ASSAY PROCEDURE in the AIA-PACK section of this analyte application. For additional procedural instructions, refer to the AIA System Operator's Manual.

1. Make the desired dilutions of the Calibration Verification Material with Sample Diluting Solution and mix well. A sufficient amount should be made to assay each diluted sample in triplicate.
2. Program the instrument to run each of the prepared dilutions three times.
3. Add the appropriate amount of each diluted sample to sample cups (refer to the instrument worklist for the amount required in each sample cup).
4. Place the sample cups and test cups on the instrument.
5. Start the assay.
6. Record the values obtained on the appropriate forms and calculate the desired statistics.

Assignment of Values

The Calibration Verification Material contains an assigned concentration of Ferritin. The assigned value is determined on a lot-to-lot basis and is designed to approximate the upper linear range of the assay. The Calibration Verification Material has been standardized against WHO 1st IS 80/602 (1983).

Results

- Make a determination of the acceptability of the data according to the specific guidelines established by your institution which satisfy the regulatory requirements under which your institution operates. If results do not meet your specifications, please initiate corrective action, as appropriate, or contact Tosoh Bioscience, Inc. for assistance.
- Retain test records in the laboratory in the designated location for future reference.

Limitations

The AIA-PACK FER Calibration Verification/Linearity Test Set is designed solely for use with AIA-PACK assay procedures. The AIA-PACK FER Calibration Verification/Linearity Test Set will only verify linearity between the established sensitivity of the assay and the value of analyte level listed on the Calibration Verification Material label.

References

1. AIA Analyte Application Manual (AAM). Tosoh Bioscience, Inc., Grove City, OH.
2. AIA-System Operator's Manual. Tosoh Bioscience, Inc., Grove City, OH.

	TOSOH CORPORATION Bioscience Division 2-2-1 Yaesu, Chuo-ku, Tokyo 104-0028 Japan Phone: +81 3 6636 3734 Fax: +81 3 6636 3627 Website: www.tosoh.com
Supplied by	TOSOH BIOSCIENCE, INC. 3600 Gantz Road, Grove City, OH 43123 USA Phone: +1 650 615 4970 Fax: +1 650 615 0415 Phone: (800) 248 6764 Fax: (800) 685 7595 Website: www.tosohbioscience.us

 European Conformity	 In vitro diagnostic medical device	 Consult instructions for use	 Temperature limitation
 Batch code / Lot number	 Manufacturer	 Authorized representative in the European Community	 Use by date
 Catalogue number / Part number	 Supplied by	 Sufficient for	