

# Certificate of Traceability

## Manufacturer Certification

This certifies that **Tosoh Corporation**, using **50 sec Analysis Mode on GR01** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:


Instrument: <b>HLC-723GR01</b>	Calibrator Lot: <b>ZB3051</b>	Column Lot: <b>X</b>
Reagent Lot: <b>XR-103X, XR-203X, XR-303X, HW-10N</b>	Calibrator Assigned Values: <b>5.78%, 10.87%</b>	<b>Ver. 1.10</b>

Date of Certification: October 1, 2025

Certification Expires: October 1, 2026



NGSP Steering Committee Chair



NGSP Network Coordinator



SRL director/ supervisor