



Certificate of Traceability

Manufacturer Certification

This certifies that **TOSOH Corporation**, using **50 sec Analysis Mode on GR01** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: HLC-723GR01	Calibrator Lot: JB5010	Column Lot: A
Reagent Lot: AR-102N, AR-203N, R-303N, HW-38N	Calibrator Assigned Values: 5.78%, 10.78%	

Date of Certification: June 1, 2026

Certification Expires: June 1, 2027

A handwritten signature in black ink, appearing to read "David Sacks", is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read "Curt Rohlfing", is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, consisting of stylized initials, is written over a horizontal line.

SRL director/ supervisor