



Certificate of Traceability

Manufacturer Certification

This certifies that **TOSOH Corporation**, using **1.6 min Analysis Mode on G8 Ver. 5.24F** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#12**

The system evaluated was:

Instrument: HLC-723G8	Calibrator Lot: JB5010	Column Lot: W
Reagent Lot: W8-101N, W8-201N, W8-301N, HW-38N	Calibrator Assigned Values: 5.78%, 10.78%	

Date of Certification: June 1, 2026

Certification Expires: June 1, 2027

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor